



CERTIFICATE OF ACCREDITATION

ANSI National Accreditation Board

11617 Coldwater Road, Fort Wayne, IN 46845 USA

This is to certify that

JAC Manufacturing, Inc.
701 Industrial Blvd.
Palmyra, WI 53156

has been assessed by ANAB and meets the requirements of international standard

ISO/IEC 17025:2017

while demonstrating technical competence in the field of

CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of activities to which this accreditation applies

L1151-1

Certificate Number


ANAB Approval

Certificate Valid Through: 11/27/2022
Version No. 002 Issued: 10/15/2019



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

JAC Manufacturing, Inc.

701 Industrial Blvd.
 Palmyra, WI 53156
 Don Miller 262-495-2141

CALIBRATION

Valid to: November 27, 2022

Certificate Number: L1151-1

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Source ¹	(0 to 22) mA	575 μ A/A + 14 μ A	Comparison performed with a Process Calibrator
DC Current – Measure ¹	(0 to 30) mA	207 μ A/A + 12 μ A	
	(30 to 110) mA	167 μ A/A + 49 μ A	
DC Voltage – Source ¹	(0 to 110) mV	100 μ V/V + 61 μ V	
	(0.11 to 1.1) V	206 μ V/V + 210 μ V	
	(1.1 to 15) V	222 μ V/V + 2.6 mV	
DC Voltage – Measure ¹	(0 to 110) mV	330 μ V/V + 35 μ V	
	(0.11 to 1.1) V	300 μ V/V + 290 μ V	
	(1.1 to 11) V	300 μ V/V + 2.9 mV	
	(11 to 110) V	531 μ V/V + 29 mV	
	(110 to 300) V	521 μ V/V + 80 mV	
Thermocouple Millivolt Simulation – Source ¹	Type E (-200 to 1 000) °C	0.9 °C	Comparisons performed with a Process Calibrator and Electronic Thermometer
	Type N (-200 to 1 300) °C	1.1 °C	
	Type J (-210 to 1 200) °C	0.9 °C	
	Type K (-200 to 1 372) °C	1 °C	
	Type T (-200 to 400) °C	1 °C	
	Type B (600 to 1 820) °C	1.5 °C	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thermocouple Millivolt Simulation – Source ¹	Type R (-20 to 1 767) °C	1.7 °C	Comparisons performed with a Process Calibrator and Electronic Thermometer
	Type S (0 to 1 760) °C	1.7 °C	
	Type C (0 to 2 316) °C	1.5 °C	
Thermocouple Millivolt Simulation – Measure ¹	Type E (-200 to 1 000) °C	1 °C	Comparisons performed with a Process Calibrator and Electronic Thermometer
	Type N (-200 to 1 300) °C	1.5 °C	
	Type J (-210 to 1 200) °C	1.1 °C	
	Type K (-200 to 1 372) °C	1.2 °C	
	Type T (-200 to 400) °C	1.1 °C	
	Type B (600 to 1 820) °C	1.8 °C	
	Type R (-20 to 1 767) °C	2 °C	
	Type S (0 to 1 760) °C	2 °C	
	Type C (0 to 2 316) °C	2.5 °C	

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Humidity Measure ¹	(10 to 90) % RH	3.4 % RH	Comparison performed with a Thermo-hygrometer
Temperature Measure ¹ (System Accuracy Test)	(-200 to 1 100) °C	1.6 °C	Comparison performed with a Process Calibrator and Thermocouple

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature Uniformity Calibration Surveys ¹ Type J Type K	(-100 to 300) °C (300 to 1 100) °C	2.3 °F	Comparisons performed with a Multi-Channel Recorder and Thermocouples per Current AMS 2750

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for all parameters, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. L1151-1.



Vice President

