



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

**JAC Manufacturing, Inc.**  
701 Industrial Blvd.  
Palmyra, WI 53156

Fulfills the requirements of

**ISO/IEC 17025:2017**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to read 'R. Douglas Leonard Jr.', is positioned above a horizontal line.

R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 27 November 2022

Certificate Number: L1151-1



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**JAC Manufacturing, Inc.**  
 701 Industrial Blvd.  
 Palmyra, WI 53156  
 Don Miller 262-495-2141

**CALIBRATION**

Valid to: **November 27, 2022**

Certificate Number: **L1151-1**

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	
DC Current – Source <sup>1</sup>	(0 to 22) mA	575 $\mu$ A/A + 14 $\mu$ A	Comparison performed with a Process Calibrator	
DC Current – Measure <sup>1</sup>	(0 to 30) mA (30 to 110) mA	207 $\mu$ A/A + 12 $\mu$ A 167 $\mu$ A/A + 49 $\mu$ A		
DC Voltage – Source <sup>1</sup>	(0 to 110) mV (0.11 to 1.1) V (1.1 to 15) V	100 $\mu$ V/V + 61 $\mu$ V 206 $\mu$ V/V + 210 $\mu$ V 222 $\mu$ V/V + 2.6 mV		
DC Voltage – Measure <sup>1</sup>	(0 to 110) mV (0.11 to 1.1) V (1.1 to 11) V (11 to 110) V (110 to 300) V	330 $\mu$ V/V + 35 $\mu$ V 300 $\mu$ V/V + 290 $\mu$ V 300 $\mu$ V/V + 2.9 mV 531 $\mu$ V/V + 29 mV 521 $\mu$ V/V + 80 mV		
Thermocouple Millivolt Simulation – Source <sup>1</sup>	Type E (-200 to 1 000) °C	0.9 °C		Comparisons performed with a Process Calibrator and Electronic Thermometer
	Type N (-200 to 1 300) °C	1.1 °C		
	Type J (-210 to 1 200) °C	0.9 °C		
	Type K (-200 to 1 372) °C	1 °C		
	Type T (-200 to 400) °C	1 °C		
	Type B (600 to 1 820) °C	1.5 °C		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thermocouple Millivolt Simulation – Source <sup>1</sup>	Type R (-20 to 1 767) °C	1.7 °C	Comparisons performed with a Process Calibrator and Electronic Thermometer
	Type S (0 to 1 760) °C	1.7 °C	
	Type C (0 to 2 316) °C	1.5 °C	
Thermocouple Millivolt Simulation – Measure <sup>1</sup>	Type E (-200 to 1 000) °C	1 °C	Comparisons performed with a Process Calibrator and Electronic Thermometer
	Type N (-200 to 1 300) °C	1.5 °C	
	Type J (-210 to 1 200) °C	1.1 °C	
	Type K (-200 to 1 372) °C	1.2 °C	
	Type T (-200 to 400) °C	1.1 °C	
	Type B (600 to 1 820) °C	1.8 °C	
	Type R (-20 to 1 767) °C	2 °C	
	Type S (0 to 1 760) °C	2 °C	
	Type C (0 to 2 316) °C	2.5 °C	

**Thermodynamic**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Humidity Measure <sup>1</sup>	(10 to 90) % RH	3.4 % RH	Comparison performed with a Thermo-hygrometer
Temperature Measure <sup>1</sup> (System Accuracy Test)	(-200 to 1 100) °C	1.6 °C	Comparison performed with a Process Calibrator and Thermocouple

**Thermodynamic**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature Uniformity Calibration Surveys <sup>1</sup> Type J Type K	(-100 to 300) °C (300 to 1 100) °C	2.3 °C	Comparisons performed with a Multi-Channel Recorder and Thermocouples per Current AMS 2750

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for all parameters, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. L1151-1.



R. Douglas Leonard Jr., VP, PILR SBU

